Product Na	ame:	ERF302WWW/0
Descrip	ption:	(Explosion Proof) Refrigerator/Freezer Combination 30 Cu. Ft. Capacity (25 Ref./5 Freezer) One solid swing door/one interior solid swing door-freezer chamber only, right hinged, not reversible Mechanical thermostat sealed in UL listed explosion proof box Adjustable operating temperature range: Refrigerator: $1^{\circ}C - 10^{\circ}C$ Freezer: $-15^{\circ}C25^{\circ}C$ Three adjustable refrigerator shelves Keyed door lock Magnetic door gasket for positive seal Stainless steel door handle Leveling legs White textured cabinet and door High density urethane foam cabinet and door insulation CFC-free refrigerant (R404a), foam insulation and packaging Exterior dimensions: $35 \frac{3}{4}$ "W x $36 \frac{1}{4}$ "D x $77 \frac{1}{2}$ "H Manual defrost Not equipped with line cord and plug; must be hardwired in conduit to 120 VAC power source 115V, 60Hz, 6 Amps, $1/3$ HP UL/C-UL listed One year parts and labor warranty Meet NFPA and OSHA guidelines for hazardous locations, Class 1, Division II, Group C & D. These units are not intended for use in general clinical or research laboratories.
		(Explosion Proof) Refrigerator/Freezer Combination 30 Cu. Ft. Capacity (25 Ref./5 Freezer) One solid swing door/one interior solid swing door-freezer chambe only, right hinged, not reversible Mechanical thermostat sealed in UL listed explosion proof box Adjustable operating temperature range: Refrigerator: $1^{\circ}C - 10^{\circ}C$ Freezer: $-15^{\circ}C25^{\circ}C$ Three adjustable refrigerator shelves Keyed door lock Magnetic door gasket for positive seal Stainless steel door handle Leveling legs White textured cabinet and door High density urethane foam cabinet and door insulation CFC-free refrigerant (R404a), foam insulation and packaging Exterior dimensions: $35 \frac{3}{4}$ "W x $36 \frac{1}{4}$ "D x $77 \frac{1}{2}$ "H Manual defrost Not equipped with line cord and plug; must be hardwired in conduit to 120 VAC power source 115V, 60Hz, 6 Amps, $1/3$ HP UL/C-UL listed One year parts and labor warranty Meet NFPA and OSHA guidelines for hazardous locations, Class 1 Division II, Group C & D. These units are not intended for use in general clinical or research